

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005)**

**RTI Electronics, Inc.**  
 33 Jacksonville Road, Bldg.1  
 Towaco, NJ, 07082  
 Stephen Honeywell 973-439-0242

**CALIBRATION**

Valid to: **November 21, 2021**

Certificate Number: **L2338**

**Ionizing Radiation**

| <b>Parameter/<br/>Equipment</b>                        | <b>Range</b>   | <b>Reference Beam Quality</b><br>(IEC 61267 code) or [PTB code] or<br>{University of Wisconsin code},<br>Target/Filtration - Reference Voltage  | <b>Expanded<br/>Uncertainty of<br/>Measurement</b><br>(+/-) | <b>Reference<br/>Standard,<br/>Method, and/or<br/>Equipment</b> |
|--|----------------|---|---|---|
| Radiography,<br>Dose Calibration<br>Factor (Gy/C)      | (35 to 150) kV | (RQR 5) W/2.83 mm Al – 70 kV<br>(RQA 5) W/23.8 mm Al – 70 kV<br><br>(RQR 9) W/3.75 mm Al – 120 kV   | 1.3 % of reading  | Radcal RC6M Ion<br>Chamber<br><br>Radcal RC6 Ion<br>Chamber     |
| Mammography,<br>Dose Calibration<br>Factor (Gy/C)      | (18 to 49) kV  | [MMV 28] Mo/0.03 mm Mo – 28 kV<br>[MRV 30] Mo/0.025 mm Rh – 30 kV<br>[WRV 30] W/0.05 mm Rh – 30 kV<br>[WSV 30] W/0.05 mm Ag – 30 kV<br>[WAV 30] W/0.5 mm Al – 30 kV<br>[WAH 30] W/2.5 mm Al – 30 kV | 1.3 % of reading  | Radcal RC6M Ion<br>Chamber                                      |
| CT Dose, Dose<br>Calibration Factor<br>(Gy/C)          | (35 to 150) kV | {UW 150 M} W/2.7 mm Al + 0.25 mm Cu<br>– 150 kV<br>{UW 120 M} W/6.7 mm Al – 120 kV<br>{UW 100 M} W/4.7 mm Al – 100 kV<br>{UW 80 M} W/2.7 mm Al – 80 kV  | 2.1 % of reading  | Standard Imaging<br>A101 Ion Chamber                            |
| CT-Dose Profiler,<br>Dose Calibration<br>Factor (Gy/C) | (35 to 150) kV | (RQR 9) W/3.75 mm Al – 120 kV   | 1.4 % of reading  | Radcal RC6 Ion<br>Chamber                                       |

**Electrical – DC/Low Frequency**

| Parameter/Equipment                | Range  | Expanded Uncertainty of Measurement (+/-)                | Reference Standard, Method, and/or Equipment                                  |
|------------------------------------|--|--|---|
| Electrometers, DC Current-Generate | 2 pA<br>(25 to 100) pA<br>100 pA to 10 $\mu$ A | 6.5 % of reading<br>0.7 % of reading<br>0.3 % of reading | Keithley 263  |
| MAS meters, DC Current-Generate    | 0.5 mA to 10 A                                 | 0.5 % of reading   | Fluke 287 & 189, Electrometer   |
| Electrometers, Charge Generate     | 2 pC<br>(2 to 100) pC<br>100 pC to 100 mC      | 6.5 % of charge<br>1.5 % of charge<br>0.19 % of charge   | Keithley 263  |
| Non-Invasive kVp Meters Measure    | (18 to 150) kV                                 | 0.7 % of reading   | HV Measurement System<br>(Radiography, Fluoroscopy, Mammography, Dental & CT) |

**Photometry and Radiometry**

| Parameter/Equipment              | Range                           | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|---------------------------------|---|--|
| Luminance Responsivity Measure   | (10 to 1 000) cd/m <sup>2</sup> | 3.3 % of reading                          | Reference Detector L100                      |
| Illuminance Responsivity Measure | (10 to 100) lux                 | 3.2 % of reading                          | Reference Detector L100                      |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. L2338.



Vice President