



Declaration of Conformity

We, RTI Group AB, Flöjelbergsg. 8 C, SE-431 37 MÖLNDAL, Sweden, declare under our sole responsibility that the product:

Product name: Piranha
Models and accessories: According to appendix 1
Type of equipment: kV- & Electrometer
Intended use of this product: According to appendix 2

is in conformity with the provisions of the following EC Directive(s):

- 2014/30/EU Electromagnetic Compatibility (EMC) Directive
- 2014/35/EU Low Voltage Directive (LVD)
- 2015/863/EU Restriction of Hazardous Substances in EEE Directive (RoHS 3)

and that the following standards and/or technical specifications referenced below have been applied:

EN 55032/B, EN 55014, EN 55024, EN 60950-1, EN 61000-3-2,
EN 61000-3-3, EN 61000-4-2, EN 61000-4-3, EN 61000-4-4,
EN 61000-4-5, EN 61000-4-6, EN 61000-4-11,
EN 62368-1 + AC1:2015, ETS 300 342-1

Möln dal, 2020-10-01

.....
Fredrik Ljungberg, President



Declaration of Conformity - Appendix 1

Piranha models

140 to 657

Accessories

<i>Probe's name</i>	<i>Part number</i>	<i>Notice</i>
RTI Dose Probe	9630003-00	s/n: 1403099 and above
RTI T20, Dose Probe	9630015-00	
RTI Light Probe	9630007-00	
RTI MAS-1	9630005-00	product version 3.0
RTI MAS-2	9630006-00	
RTI CT Dose Profiler	9630013-00	product version 2.0
RTI Chamber Adapter	9630016-00	product version 1.1
RTI CT Ion Chamber 10 cm	9630025-00	
RTI CT Ion Chamber 30 cm	9630026-00	
CT chamber DCT-10	9605111-00	
CT chamber DCT-30	9605135-00	
Ion Chamber Magna 1cc	9606100	
DAP Chamber 147x147 mm	9605060-00	
DAP Chamber 86x86 mm	9605070-00	
GE mA Test Point Cable	9630096-01	



Declaration of Conformity - Appendix 2

Intended use

Together with external probes the Piranha System is intended to be used for independent service and quality control, including measurements of kerma, kerma rate, kVp, tube current, exposure time, luminance, illuminance, and dose area product, within limitations stated below.

When installed according to accompanying documents, the product is intended to be used together with all diagnostic X-ray equipment except for:

- therapeutical X-ray sources.
- X-ray equipment with tube potential below 18 kV or above 160 kV.
- X-ray equipment on which the instrument cannot be mounted properly.
- specific types of X-ray equipment listed in the instructions for use or in additional information from the manufacturer.

With the X-ray installation without patient present, the product is intended to be used:

- for assessing the performance of the X-ray equipment.
- for evaluation of examination techniques and procedures.
- for service and maintenance of the X-ray equipment.
- for quality control of the X-ray equipment.
- for educational purposes, authority supervision etc.

The product is intended to be used by hospital physicists, X-ray engineers, manufacturer's service teams, and other professionals with similar tasks and competencies. The operator needs training to be able to use the product as intended. This training can be achieved either by study of the manual, study of the built-in help functions in measurement software or, on request, by a course ordered from the manufacturer.

The product is intended to be used inside X-ray rooms ready for clinical use and can safely be left switched on and in any measuring mode in the vicinity of patients.

The product is **NOT** intended to be used:

- for direct control of diagnostic X-ray equipment performance during irradiation of a patient.
- so that patients or other unqualified persons can change settings of operating parameters during, immediately before, or after measurements.
- for any guidance to diagnosis of patients.